

1.3cm secondary focus receiver

(21600-24400 MHz)

This is a 2-channel system for VLBI observations.
In fall 2013 this system will be replaced by a new double-horn receiver.

Calibration Information

Frequency [GHz]	Channel	Polarisation	Tcal [K]	Tsys [K]	Sensitivity [K/Jy]	SEFD [Jy]	Aperture Eff. [%]	TMB/S [K/Jy]	Main Beam Eff. [%]	FWHM [arcsec]	Last update
21.73	A	LCP	5.1	78	0.90	96	28	1.8	47	36.5	Jul 2007
21.73	B	RCP	4.9	84	0.90	105	28	1.8	47	36.4	Jul 2007
22.23	A	LCP	5.0	91	0.9	110	30	1.8	47	36.5	Jul 2008
22.23	B	RCP	5.0	88	0.9	105	30	1.8	47	36.5	Jul 2008
23.05	A	LCP	3.4	83	0.9	97	30	1.8	48	36	Jul 2008
23.05	B	RCP	3.7	70	0.9	91	30	1.8	47	36	Jul 2008
normalized Gain curve (G = A0 + A1·Elv + A2·Elv2)							Observed in	confirmed			
A0 = 0.95282		A1 = 3.6092E-3		A2 = -6.8913E-5		Mar 2003					
A0 = 0.91119		A1 = 4.7557E-3		A2 = -6.2902E-5		Feb 2007					

Comments:

- Note that the system temperature at this frequency depends strongly on weather conditions!
- The new gain curve (Feb 2007) was corrected for opacity. Because of the previous point it is always recommended to correct data at 1.3cm for opacity.
- Zenith tau values can range from 0.05 to 0.2.

Version description for OBSINP

RX Name	Wavelength [cm]	Frequency (center) [GHz]	Nr. of Horns
S13mm Multifrequency RX	1.3	21.7-24.4 (23.05)	1
Version:	Comment		
1. Continuum (BW: 2 GHz)	Broad Band Continuum		
2. Line (BW: 100 MHz)	Spectroscopy/Continuum using narrow band IF + VLBI IF Polarimeter		
3. Line (BW: 500 MHz)	Spectroscopy/Continuum using VLBA IF + VLBA IF Polarimeter		
4. Pulsar (BW: 500 MHz)	Pulsar 500 MHz BW Version		
Horn offsets [arcsec]	-593.0,55.4		

Channel assignment in the MBFITS data files

Note that the narrow line and VLBA IF channels are usually only available when the specific line version of the receiver was selected. In addition for most receivers with narrow line channels the cables at the patch board need to be connected by the receiver group.

To select different channel numbers in OBSINP, the online plot, or the toolbox the numbers have to be specified like $c(1)+c(2)$ to add channel 1 and 2. E.g. channel 1 and 2 contain the LCP and RCP broadband channels, then "OnlPlot pen='c(1)+c(2)'" or "toolbox use='c(1)+c(2)'" will select these channels. In OBSINP the pen can be directly specified in the receiver selection menu.

Abbreviations:

SB: narrow band channel (Schmalband-Kanal), 100 MHz band width

BB: digital broad band channel (Breitband-Kanal), band width varies for different receivers

VLBA: VLBA IF, 500 MHz band width

BW: band width

TP: total power

2cm SFK single horn receiver			
Channel	IF	Pol.	Comment
1	SB	LCP	TP A
2	SB	RCP	TP B
3	SB	cross	cos AB
4	SB	cross	sin AB
5	VLBA	LCP	TP A
6	VLBA	RCP	TP B
7	VLBA	cross	cos AB
8	VLBA	cross	sin AB
9	BB	LCP	TP A
10	BB	RCP	TP B

Tcal measurements

VLBI calibration was only performed for the lower part of the band. Frequencies above 22.8 GHz will follow.



From:
<https://eff100mwiki.mpifr-bonn.mpg.de/> - Effelsberg 100m Teleskop

Permanent link:
https://eff100mwiki.mpifr-bonn.mpg.de/doku.php?id=information_for_astronomers:rx:s13mm&rev=1371737860

Last update: 2013/09/30 13:57

